

LCRT-2005-S

Product tags: VIS



Description

Light transmission

Light transmission is the visually sensed light permeability of materials. It is particularly important when it comes to specification of window panes on all types of vehicles, display sheets and disks as well as all the other samples whose transmission is assessed with the photometric responsivity $(V(\lambda))$ of the human eye.

Light transmission measurement

Light transmission is a relative measurand. It describes the signal difference of the measured luminance in a geometrically defined radiation beam path. The measurements are done both without a test sample (100%) and with a test sample. The light source of the measurement device must be characterized by a spectrum that corresponds to the standard illuminant type A, Cor D50. The spectral responsivity of the receiver must match that of the human eye. These properties allow ECER43 conform measurements.

LCRT-2005-S light transmission measurement device

The LCRT-2005-S is elaborately designed for light transmission measurement of thin, scratched and clear samples. The D/O measurement geometry comprises of an integrating sphere light source and a luminance measurement device. The monitor detector of the light source and that of the receiver are equipped with a diode array. The spectral measurement data enables precise simulation of the standard light spectra and the photometric responsivity spectrum of the receiver. The sample is aligned in front of the light source for measurement. The degree of light transmission can thus be determined through diffuse sample illumination for thin scratched samples as well.


LCRT-2005-S spectrophotometer

The LCRT-2005-S can also be used as a spectrophotometer due to its spectral measurement detectors in source and receiver. The spectral range is between 425 and 705nm.


Measurement device for freehand measurements

The integrating sphere light source is well guarded against shock and stains by its synthetic coating, LED lamps and the protective glass on the illumination field. In order to minimize any effects by ambient light, the measurement is performed using pulsed light. The device is also equipped with a camera to aid in alignment of the source and receiver for freehand measurements. The device can be powered using four AA batteries or via USB. A hard-top casing is also supplied for safe storage and transport of






Measurement geometry: 100% setup



Measurement geometry for thin, non-scattering samples



Measurement geometry for thin, scattering samples

the device as well as its spare batteries and accessories.

Fast and simple freehand measurements

One of the characteristic properties of the LCRT-2005-S is its ability to easily perform a quick measurement in only a couple of minutes:


- 1) Connection of the source and receiver to the controller device
- 2) 100% adjustment measurement
- 3) Alignment to the test sample
- 4) Automatic start of the measurement
- 5) Display of the measurement values

USB interface with readout software

The USB interface enables data readout and power supply. The software delivered with the device can be used for measurement data readout.

Traceable measurements

Two transmission standards with 70% and 80% light transmittance are offered for device matching in applications where traceable measurement values are required.



Freehand light transmission measurement: 1) 100% adjustment 2) sample alignment (DUT) 3) automatic measurement start upon setup, 4) Display of the measurement values

Specifications

| General | | | |
|----------------------|---|--|--|
| Short description | Mobile measuring Instrument for light transmission. | | |
| Main features | Spectral measuring method. Compact source and receiver. Build in camera for receiver to source alignment support. Controller with battery (four AA) or USB power operation. Puls modulated LED lamp for measurements in ambient light conditions. | | |
| Measurement range | Spectral range: 425 nm to 705 nm Spectral resolution: 5 nm Transmission range: 5 % to 100 % Measurement beam diameter: 6.6 mm Illumination: A, C, D65 Detection: Spectral Photometric, Spectral Radiometric | | |
| typical applications | Measurement of the light Iransmission of automotive and all other vehicles front and side windows. Measurement of the spectral transmission of thin films and thin plates. | | |
| Calibration | Realative measurement methode with 100 % reference adjustment. Optional calibration with 70 % or 80 % calibration standards. | | |
| Product | | | |
| Measurement geometry | D/0 geometry; measurement of the luminance ratio with a diffuse light source as per CIE 130 & DIN 5036. | | |

| Beam diameter | 6.6 mm at contact measurement | | |
|---------------------------------|--|--|--|
| Light Source | Integrating sphere light source with a 20 mm illumination field; LED lamps; diode array monitor detector. | | |
| | Simulation of the standard illuminants type A, D65 and D50. | | |
| Sensor | Diode array detector with radiance lens. Depolarizer for measurement of polarized samples. Simulation of photometric responsivity. Digital camera for aid in freehand setup of source and receiver. | | |
| spectral range | 425 nm to 705 nm | | |
| Measurement range | 5 % to 100 % transmission for a color-neutral transmission spectrum | | |
| typical Measurement uncertainty | ± 1 % absolute | | |
| Data Resolution | 0.1 % | | |
| Calibration | Relative measurements after performing 100% match against air. Traceable measurements after matching with calibrated standard filters. | | |
| Source | | | |
| Light Source | White LEDs in pulse mode, usable wavelength range: 425 nm to 705 nm | | |
| Monitor detector | array spectrometer based on a 256 pixel diode | | |
| Connector | Length 1.5m | | |
| | Mini DIN plug | | |
| | RS232 and voltage supply | | |
| Housing | Aluminium profile with plastic caps | | |
| | Threaded bores for mounting | | |
| Dimensions | 160 mm x 45 (60) mm x 85 mm | | |
| Weight | 450 g | | |
| Light Source | Integrating sphere with synthetic ODM98 coating. 20 mm illumination field diameter with homogenous luminance distribution (lambertian radiator). Illumination field equipped with a protective shield. | | |
| Receiver | | | |
| Sensor | 256 pixel diode array spectrometer with an achromatically corrected lens. Implemented depolarizer for measurement of polarized samples. | | |
| Measurement beam geometry | Measurement field angle - 0.38 $^{\circ}$ | | |
| | Sample alignment 0 ° | | |
| | illumination field diameter by contact measurement - 6.6 mm, in 1 m measurement distance 12.6 mm | | |
| Dimensions | 160 mm x 45 mm x 85 mm | | |
| Weight | 400 g | | |
| Controller | | | |
| Source and receiver connector | Two mini DIN plug connections | | |
| Display | Monochromatic display with backlight that can be switched on/off | | |
| Parameter adjustment | Menu operated | | |
| | saving of the last used settings | | |
| | four control buttons | | |
| | | | |

| Interface | USB | |
|-------------------|---|--|
| Power Supply | 4 x AA batteries alternative 4 x AA batteries with external charger USB | |
| Dimensions | 230 mm x 72 (115) mm x 35 (50) mm | |
| Weight | 400 g | |
| Miscellaneous | | |
| temperature range | 10 °C to 40 °C | |
| Humidity | Above the saturation temperature | |

Configurable with

| Product Name | Product Image | Description |
|---------------------|---------------|--|
| B2S-40-TRTH | | Rail-bench to extend the LCRT-2005-S use to evaluate thin samples transmission by diffuse and regular illumination. Features: stable 1m length rail bench with stand for source and receiver. Carriage with sample holder. |
| PMS-RIT | | Stand to extend the LCRT-2005-S use to measure the regular (inline) transmission of thick samples up to 100mm thickness. Features: stable stand with mount for source and receiver. Sample table. |

Purchasing information

| Article-No. | Model | Description |
|-------------|-------------|--|
| Product | | |
| 15296702 | LCRT-2005-S | Measurement device, 100% adjustment support plate, setup aid, hard case, software CD, handbook |

Software

| Article-No. | Model | Description |
|-------------|---------------------|--|
| 15312082 | S-SDK-LCRT2005 | Software development kit for the implementation of an LCRT-2005-S device into custom software. |
| Accessories | | |
| 15297874 | LCRT-2005-S-BN-T70 | Spectral calibration standard with 70% light transmission; calibration certificate |
| 15297875 | LCRT-2005-S-BN-T80 | Spectral calibration standard with 80% light transmission; calibration certificate |
| 15305907 | LCRT-2005-S-BN-T100 | 100% alignment plate |
| 15298554 | LCRT-2005-S-Z01 | Bench-top stand for source and receiver |
| 15298640 | B2S-40-TRTH | Optical bench with adjustable sample holder |
| 15297916 | PMS-RIT | Bench-top stand for source and receiver |

